

[Cedrus](#)[Products](#)[Ordering](#)[Support](#)[XID Home](#)[Device Properties](#)[Timing Features](#)[Supported Protocols](#)[Accessory Connector](#)[Voice Key](#)[Test Procedures](#)[Misc. Commands](#)

XID Accessory Connector Command

Re-Program Flash Memory (f3)

What it Does

Instructs the XID device to get ready to receive an update to itself.

This command does not work on **SV-1**.

Using This Command

Do not use this command!

If you do by any chance, the XID device will most likely hang. Try powering it off, wait a few seconds, and then re-start it.

Corresponding Inquiry Command

This command has no inquiry equivalent.

© Copyright 2014 **Cedrus Corporation**, P.O. Box 6309, San Pedro, CA 90734 - USA

Toll Free: (800) 233-7871, other: (310) 548-9595. Send us [an email](#) or join [mailing list](#). 